


UHPLC CERTIFICATE OF ANALYSIS

SAMPLE INFORMATION


Sample received:

Product Name	NAD+ 500mg Batch No. 50106
Sequence	Nicotinamide-ribose-5'-phosphate – pyrophosphate – adenosine-ribose-5'-phosphate
Dissolution Condition	100% UHPLC water with 0.1% Trifluoroacetic Acid (TFA). Dissolved sample picture included here. 
Molecular Weight	~663.43 g/mol

STORAGE CONDITION: 0 °C

SAMPLE NOTES

One 3 ml clear glass vial with silver cap and silver aluminum crimp. Lyophilized powder present.



PUMP SETTINGS

Pump A	100% H2O with .1% Trifluoroacetic Acid
Pump D	Acetonitrile + 0.1% FA
Column Usage	Agilent Zorbax 300SB-C18, 2.1 x 100 mm, 1.7 µm, 300 Å
Gradient	<ul style="list-style-type: none">- 2% D hold (0–0.5 min)- 2 → 15% D over 6 min (≈2.17% D/min)- 15 → 30% D over 3 min- Hold 30% D (2 min)- Re-equilibrate to 2% D over 3.5 min- Total run time: ~15 min.
Pump Settings	0.6 mL/min
Injection Volume:	10 µL
Temperature	30 °C


PURITY RESULT

Purity is **99.82%**

Product weight extrapolation: 509.1mg

Total impurity 0.18% (trace inert solvents and minor unidentified peaks; no significant related substances or degradation products detected above 0.1%).

Chromatogram Export: please note that x axis represents elution window.



CONCLUSION

This sample was analyzed on a Thermo Ultimate U3000 HPLC stack using reverse-phase high-performance liquid chromatography and determined to contain **99.82% NAD+**, with the remainder being impurities of minor significance.

CERTIFICATION

Certified by: Jonathan Barber

Title: Analytical Chemist

Date: 15 September 2025



Trulab Peptides www.trulabpeptides.com | **Phone:** 872-239-6582

Address: 15017 North Dale Mabry Highway #1121 Tampa, FL 33618 United States